

## **What products are not affected and why?**

Products that are not affected may have different sound abatement foam materials, as new materials and technologies are available over time. Also, sound abatement foam in unaffected devices may be placed in a different location due to device design.

**Products not affected by this recall notification (U.S. only) / field safety notice (International Markets) include:**

- Trilogy Evo
- Trilogy Evo OBM
- Trilogy EV300
- Trilogy 202
- BiPAP A40 EFL
- BiPAP A40 Pro
- M-Series
- DreamStation 2
- Omnilab (original based on Harmony 2)
- Dorma 100, Dorma 200, & REMStar SE
- All oxygen concentrators, respiratory drug delivery products, airway clearance products.